

**OFFICE OF THE DIRECTOR OF AGRICULTURE & FOOD PRODUCTION, ODISHA,
KRUSHIBHAWAN, BHUBANESWAR.**

4P(07)28/2018 **24185**

/Agril date **05.09.2019**

CORRIGENDUM

In partial modification of the tender call notice no. 4P(07)28/2018-22815 /Agril dt. 19.08.2019 published in THE NEW INDIAN EXPRESS on 20.08.2019, PRAGATIBADI and PRATIDINA, the Odia daily news paper on 21.08.2019 for procurement of Mobile Soil Testing Laboratory (Composite Unit), Atomic Absorption Spectrophotometer, High Performance Liquid Chromatograph and Gas Chromatograph for Soil and Insecticide Testing Laboratories of the state the following corrigendum has been made:

1. Annexure –II, Set-D, Equipments for Insecticide Testing Laboratory-High Performance Liquid Chromatography- Point no.4- Column Oven : “Should have a range of ambient 10⁰ C to 60⁰ C or better” may be read as “**Should have a range 10⁰ C below ambient to 60⁰ C or better**”.
2. Annexure –II, Set-D, Equipments for Insecticide Testing Laboratory-High Performance Liquid Chromatography- Point no. 5 UV-Vis Detector: “Should have D2 lamp light source” may be read as “**Should have D2 and Tungsten lamp light source** ”

Sd/-

(Dr. M. Muthu Kumar)
Director of Agriculture &
Food Production, Odisha